

CERTIFICATE OF ANALYSIS

Page 1 of 1

Product Description: MSD[®] Apoptosis Whole Cell Lysate Kit **Catalog Number:** K15102D - Series

Lot Number: K0013901

Expiration Date: N/A (See expiration date on specific components below)

Statement: This product was manufactured and tested according to approved MSD documents.

This lot meets MSD's specifications.

Kit Contents:

Plate					
Lot #	Z0055669	Catalog #	N45102B	Expiration Date:	31 DEC 2018

Antibody Blend		
Lot Number:	D0013500	
Catalog Number:	D20DE	
Expiration Date:	28 FEB 2019	
Storage Temperature:	2-8°C	

Antibody Blend		
Lot Number:	D0013501	
Catalog Number:	D21DB	
Expiration Date:	28 FEB 2019	
Storage Temperature:	2-8°C	

Antibody Blend		
Lot Number:	D0013502	
Catalog Number:	D21CF	
Expiration Date:	28 FEB 2019	
Storage Temperature:	2-8°C	

Specifications:

Release method:

Coating Uniformity	Specifications	
Mean Intraplate CV	≤ 15%	
Mean Interplate CV	≤ 18%	
No more than one plate with CV	> 18%	
No plate with CV	> 20%	

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Vanessa Castleberry	Quality Assurance	Vanessa Castlebury	16JUN2017

COA-8384-A QA-FM-153-A