

## CERTIFICATE OF ANALYSIS

COA - 04102

**Product Description:** MA2400 384 Plate Pack

**Catalog** #/**Size:** L25XA-1 / 1 Plate; L25XA-4 / 15 Plates; L25XA-6 / 120 Plates; L25XA-7 / 510 Plates

Lot Number: 11D01A

**Expiration Date:** 30 JUN 2015

Statement: This product was manufactured and tested according to approved MSD documents. This

lot meets MSD's specifications.

BON (Electrical resistivity): < 200 microohms in all circuits

Leak test: No leaks Free Tag 15 000 test

ECL (mean  $\pm$  tolerance):  $18\,000 \pm 3,600$  cts Max Average Intraplate CV for any spot: 5%Maximum Intraplate CV for any plate: 9%Max Interplate CV for any spot: 8%

## The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Preparer	Rolanda Spratley	Biologics/Chemistry	Bu July Fally	28 JAN 2014
Review/Approval	Michael Eager	Quality Control	791	10 FEB 2014