

## CERTIFICATE OF ANALYSIS

COA - 03445

**Product Description:** MULTI-ARRAY® 96-well High Bind Plate Pack

Catalog #/Size: L15XB-1 / 1 Plate; L15XB-3 / 10 Plates; L15XB-6 / 100 Plates; L15XB-7 / 500 Plates

**Lot Number:** 22C0CA **Expiration Date:** June 2014

Statement: This product was manufactured and tested according to approved MSD documents. This

lot meets MSD's specifications.

BON (Electrical resistivity): < 200 microohms in all circuits

Leak test: No leaks Free Tag 15 000 test

ECL (mean  $\pm$  tolerance): Pass

Max Average Intraplate CV for any spot: 3% Maximum Intraplate CV for any plate: 7%

Max Interplate CV for any spot: 7%

## The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Preparer	Rolanda Spratley	Biologics/Chemistry	(2 m)	15 Spu13
Review/Approval	Michael Eager	Quality Control	9/110	15 MV2013
			11/0	, ,,,