

CERTIFICATE OF ANALYSIS

COA - 03870

Product Description: MULTI-ARRAY® 96-well High Bind Plate Pack

Catalog #/Size: L15XB-1 / 1 Plate; L15XB-3 / 10 Plates; L15XB-6 / 100 Plates; L15XB-7 / 500 Plates

Lot Number: 22D0GA

Expiration Date: 30 JUN 2015

Statement: This product was manufactured and tested according to approved MSD documents. This

lot meets MSD's specifications.

BON (Electrical resistivity): < 200 microohms in all circuits

Leak test: No leaks Free Tag 15 000 test

ECL (mean \pm tolerance): 27 400 \pm 5,480 cts Max Average Intraplate CV for any spot: 3% Maximum Intraplate CV for any plate: 7%

Max Interplate CV for any spot: 7%

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Preparer	Rolanda Spratley	Biologics/Chemistry	Taller	2358P2013
Review/Approval	Michael Eager	Quality Control	75N	034672013
			///	, , ,