

CERTIFICATE OF ANALYSIS

COA - 06876

Product Description: MULTI-ARRAY® 96-well High Bind Plate Pack

Catalog #/**Size:** L15XB-1 / 1 Plate; L15XB-3 / 10 Plates; L15XB-6 / 100 Plates; L15XB-7 / 500 Plates

Lot Number: 22F0GA

Expiration Date: 31 AUG 2017

Statement: This product was manufactured and tested according to approved MSD documents. This

lot meets MSD's specifications.

BON (Electrical resistivity): < 200 microohms in all circuits

Leak test: No leaks Free Tag 15 000 test

ECL (mean \pm tolerance): 23 461 \pm 4, 692 cts Max Average Intraplate CV for any spot: 3% Maximum Intraplate CV for any plate: 7% Max Interplate CV for any spot: 7%

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Preparer	Rolanda Spratley	Biologics/Chemistry	Bo for fritter	09 MAY 2016
Review/Approval	Dawn Stollar	Quality Assurance	Dawn Stollar	09 MAY 2016