

CERTIFICATE OF ANALYSIS

COA – 06696

Product Description: MULTI-ARRAY® 96-well High Bind Plate Pack

Catalog #/Size: L15XB-1 / 1 Plate; L15XB-3 / 10 Plates; L15XB-6 / 100 Plates; L15XB-7 / 500 Plates

Lot Number: 22G03A

Expiration Date: 28 FEB 2018

Statement: This product was manufactured and tested according to approved MSD documents. This

lot meets MSD's specifications.

BON (Electrical resistivity): < 200 microohms in all circuits

Leak test: No leaks Free Tag 15 000 test

ECL (mean \pm tolerance): 23 461 \pm 4, 692 cts Max Average Intraplate CV for any spot: 3% Maximum Intraplate CV for any plate: 7%

Max Interplate CV for any spot: 7%

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

| | Name | Function | Signature | Date |
|-----------------|------------------|---------------------|----------------|-------------|
| Preparer | Rolanda Spratley | Biologics/Chemistry | Bu Jacky Fally | 28 MAR 2016 |
| Review/Approval | Dawn Stollar | Quality Assurance | Dawn Stollar | 28 MAR 2016 |