R-PLEX[™] Antibody Set

Human GLP-1 (inactive)

Cat. No: F201J-3, 5-plate size

F201J-8, 50-plate size

Ordering Information

MSD Customer Service Phone: 1-240-314-2795 Fax: 1-301-990-2776 Email: CustomerService@ mesoscale.com www.mesoscale.com/support

Scientific Support

Phone: 1-240-314-2798 Email: ScientificSupport@ mesoscale.com www.mesoscale.com/support

Company Address

MESO SCALE DISCOVERY® A division of Meso Scale Diagnostics, LLC. 1601 Research Boulevard Rockville, MD 20850-3173 USA

www.mesoscale.com®

For Research Use Only. Not for use in diagnostic procedures.

Contents and	Description	No. of vials	Vol. /vial	Storage
Storage	Human GLP-1 (inactive) Calibrator, 20X	5	50 μL	≤-70°C
	Biotin Human GLP-1 (inactive) Capture Antibody	1	1,250 μL	2-8°C
	SULFO-TAG™ Human GLP-1 (inactive) Detection Antibody, 100X	1	375 μL	2-8°C
	Note: Calibrators and antibodies are shipped separately.			


R-PLEX Antibody Sets provide an expanded menu of electrochemiluminescence assays for biomarker discovery and development. The sets include a matched biotinylated capture and SULFO-TAG conjugated detection antibody pair and calibrator for the quick and easy development of highly sensitive immunoassays on a variety of MSD® plates including U-PLEX® MULTI-SPOT plates and MSD GOLD™ Small Spot Streptavidin plates.

Glucagon-like Peptide 1 (GLP-1) plays a key role in the promotion of glucose-dependent insulin secretion and insulin biosynthesis. GLP-1 works in concert with insulin to inhibit glucose secretion and lower overall blood glucose levels. It plays roles in gastric emptying and in the regulation of short-term feeding behavior. The active isoforms are rapidly cleaved and inactivated by DPPIV.

Assay Performance

R-PLEX Antibody Sets can be easily adapted to develop singleplex and multiplex assays on the MSD platform. MSD GOLD Small Spot Streptavidin plates can be used for singleplex assay development, whereas multiplex assay formats can be performed using U-PLEX MULTI-SPOT® plates.

Representative data from the R-PLEX Human GLP-1 (inactive) Antibody Set tested on MSD's U-PLEX platform using Diluent 12 as assay diluent and Diluent 12 as antibody diluent is presented below. The lower limit of detection (LLOD) is a calculated concentration corresponding to the signal 2.5 standard deviations above the background (zero calibrator). The performance of your assay may vary depending on the diluent used, the sample type, and the other analytes when testing in a multiplexed format.



Top-of-Curve Conc. (pg/mL)	LLOD (pg/mL)
10,000	3.2

For developing individual assays using R-PLEX Antibody Sets, please refer to the R-PLEX Antibody Sets Singleplex Assays product insert. For multiplex assay development using R-PLEX Antibody Sets or a combination of R-PLEX and U-PLEX Antibody Sets, refer to the U-PLEX Development Pack product insert. The detailed instruction manuals as well as the R-PLEX assay datasheets are available at www.mesoscale.com/R-PLEX-documents.

Sample Dilution

The R-PLEX Human GLP-1 (inactive) assay has been tested with the following sample type(s) at the specified dilution in assay diluent.

Sample Type	Recommended Fold Dilution	
Serum	4	
Plasma	4	

The assay may be used for other sample types. We recommend performing a dilution series and testing for diluent compatibility when measuring a different sample type.





R-PLEX Antibody Set

Source

Calibrator	Recombinant Inactive Glucagon-like peptide-1	
Antibodies	Capture antibody - mouse monoclonal, detection antibody - mouse monoclonal	
Assay generation	A	

Recommended Assay Components

The following items are recommended for developing assays using R-PLEX Antibody Sets.

Plates for Singleplex Assay Development: All plates are available in 1, 5, 30, 120, and 510 plate sizes.

Description	Catalog No.	
MSD GOLD 96-well Small Spot Streptavidin Plates	L45SA	

Plates for Multiplex Assay Development: The U-PLEX platform allows you to create your own multiplex assays. Using two simple tools, a 10-spot U-PLEX plate and unique Linkers, you can build custom multiplex panels for any combination of analytes. U-PLEX Development Pack includes U-PLEX Plates, Linkers, Stop Solution, and Read Buffer T. All development packs are available in 1, 5, and 25-plate pack sizes. You can choose the appropriate Development Pack according to the number of assays you plan to multiplex, e.g., use a U-PLEX 4-Assay Development Pack if you are developing a 4-plex assay.

Description	Catalog No.
U-PLEX 2-Assay Development Pack	K15227N
U-PLEX 3-Assay Development Pack	K15228N
U-PLEX 4-Assay Development Pack	K15229N
U-PLEX 5-Assay Development Pack	K15230N
U-PLEX 6-Assay Development Pack	K15231N

Description	Catalog No.
U-PLEX 7-Assay Development Pack	K15232N
U-PLEX 8-Assay Development Pack	K15233N
U-PLEX 9-Assay Development Pack	K15234N
U-PLEX 10-Assay Development Pack	K15235N

Note: R-PLEX assays can be combined with U-PLEX assays to build multiplex panels. View the R-PLEX Assay Designer at www.mesoscale.com/R-PLEX to customize your multiplex panel.

Buffers:

Description	Catalog No.
MSD GOLD Read Buffer (200 mL)	R92TG-1
MSD Wash Buffer (20X, 100 mL)	R61AA-1

Diluent Compatibility

The R-PLEX Human GLP-1 (inactive) Antibody Set has been tested with the following diluents.

Diluent Type	Diluent Name	Catalog No.	Size
Access Diluont	Diluent 12	R50JA-4	10 mL
Assay Diluent		R50JA-3	50 mL
Antibody Dilyont	Diluent 12	R50JA-4	10 mL
Antibody Diluent		R50JA-3	50 mL

To run five plates, 50 mL of assay diluent and 40 mL of antibody diluent are required. Additional assay diluent may be required for sample dilution. Testing of different diluents can help optimize assays for specific experimental conditions. MSD offers a range of assay and antibody diluents for separate purchase. Please view our website at www.mesoscale.com/R-PLEX for more details on compatible diluents for R-PLEX assays.

The R-PLEX Human GLP-1 (inactive) Antibody Set is for research use only.





